

CERTIFICATE NO.

PL-26-00112

VERIFICATION KEY

3TW8THNN

Scan or enter both at
prymalab.net/testing/verify

CERTIFICATE OF ANALYSIS · LOT RELEASE · ISSUED FOR PURAVIUM

Tirzepatide 30 mg

Lot PL-2609-112 · Lyophilized powder · Label claim 30 mg/vial · Distributed as GIP2-TZ

TESTED IN-HOUSE BY PRYMALAB

In-house analytical result,
not an independent third-party test

LOT / BATCH NUMBER	DATE MADE	DATE OF ANALYSIS	DATE ISSUED
PL-2609-112	2026-09-18	2026-09-19	2026-10-03
RETEST / EXPIRY	UNITS IN LOT / SAMPLED	STORAGE	CERTIFICATE TYPE
2028-09-18	350 / 3	-20 °C, desiccated, dark	Lot COA, 2 pages

COMPOUND	CAS NUMBER	MOLECULAR FORMULA	THEORETICAL MASS (AVG.)
Tirzepatide	2023788-19-2	C ₂₂₅ H ₃₄₈ N ₄₈ O ₆₈	4813.5 Da
LY3298176; GIP/GLP-1 receptor dual agonist			
SEQUENCE			
H-Y-Aib-EGTFTSDYSI-Aib-LDKIAQK(2×AEEA-γGlu-C20 diacid)AFVQWLIAGGPSSGAPPPS-NH ₂ (39 residues; Aib2 and Aib13; Lys20 acylated with eicosanedioic acid via γ-Glu and two AEEA spacers; C-terminal amide)			

PURITY · RP-HPLC, AREA %

99.26%

Specification 99.0% • main peak at 8.69 min

IDENTITY · OBSERVED MASS

4813.4 Da

Theoretical 4813.5 Da • -0.1 Da (-21 ppm)

OVERALL CONCLUSION


PASS

Meets every specification for the tests performed

TEST	METHOD	SPECIFICATION	RESULT	CONCLUSION
Appearance	Visual inspection	White to off-white lyophilized powder	White lyophilized powder	PASS
Identity (mass)	LC-MS, ESI positive	Theoretical ± 1.0 Da	4813.4 Da	PASS
Purity	RP-HPLC, UV 214 nm, area%	99.0%	99.26%	PASS
Largest impurity	RP-HPLC, area%	1.0%	0.21% at 7.81 min (Trp 25 oxidation)	PASS
Water content	Karl Fischer	5.0%	1.7%	PASS
Bacterial endotoxin	LAL, kinetic	≤ 5.0 EU/mg	0.06 EU/mg	PASS
Sterility / bioburden	USP <71>, membrane filtration	No growth after 14 days	No growth; conforms	PASS

1 HPLC CHROMATOGRAM

Data file: TZP_PL-2609-112_20260919_01 · Column C18, 150 × 4.6 mm · UV 214 nm



2

HPLC METHOD · TZP_PL-2609-112_20260919_01

INSTRUMENT	DATA SYSTEM	DETECTOR
Waters Alliance e2695, SN E2695B742	Empower 3, Build 3471	Waters 2998 PDA
DETECTION WAVELENGTH	COLUMN	COLUMN PART / SERIAL
214 nm	C18, 150 × 4.6 mm, 3.5 µm	186003326 / USHX20400
MOBILE PHASE A	MOBILE PHASE B	FLOW RATE
0.1% TFA in water	0.1% TFA in MeCN	1.0 mL/min
COLUMN TEMPERATURE	INJECTION VOLUME	RUN TIME
40 °C	20 µL	12 min
SAMPLE DILUENT	SAMPLE CONCENTRATION	BLANK INJECTED
0.1% TFA in 30% MeCN	0.5 mg/mL	Yes
REFERENCE STANDARD	In-house Tirzepatide reference, lot PL-RS-TZP-01, RT 8.66 min	

3

GRADIENT PROGRAM

TIME (MIN)	% A	% B
0.0	65	35
2.0	65	35
9.0	25	75
10.0	15	85
11.0	15	85
12.0	65	35

4

PEAK TABLE

PEAK	RT (MIN)	AREA	AREA %
1	6.08	7,701	0.16
2 (deamid.)	7.60	9,145	0.19
3 (Trp-ox)	7.81	10,108	0.21
4	8.69	4,777,782	99.26
5	9.04	8,664	0.18
Total		4,813,400	100.00

Threshold 0.05% area · Blank subtracted: Y · Purity = main peak area ÷ total integrated area × 100, to two decimals; main peak shaded.

5

IDENTITY BY MASS SPECTROMETRY

INSTRUMENT	IONIZATION / MODE	OBSERVED M/Z (CHARGE)
Waters Acquity SQD	ESI, positive	963.7 (5+ [M+5H]) · 803.2 (6+ [M+6H])
DECONVOLUTED MASS	THEORETICAL MASS	DIFFERENCE
4813.4 Da	4813.5 Da	-0.1 Da (-21 ppm)
DATA FILE	ACCEPTANCE	
TZP_PL-2609-112_MS_20260919_01	Within ± 1.0 Da	

6

ADDITIONAL TESTS

TEST	PERFORMED BY	DATE	REPORT / FILE
Water content	PrymaLab	2026-09-19	RPT-KF-PL-2609-112
Endotoxin	PrymaLab	2026-09-19	RPT-ET-PL-2609-112
Sterility	PrymaLab	2026-09-19	RPT-ST-PL-2609-112
Peptide content	PrymaLab	2026-09-19	RPT-AA-PL-2609-112

A test with no row here was not performed on this lot. Outside-laboratory reports, when used, are attached and named below.

7

REVIEW AND APPROVAL

D. Reynolds	S. Calloway	M. Phelps
Analyst · performed the testing	Reviewer · checked against raw data	Lab Director · approved for release
Date: 2026-09-19	Date: 2026-10-02	Date: 2026-10-03

Raw data retained at: Empower 3 (Build 3471), project PL-2609, archive PL-ARC-2026; 7 years. Outside laboratory reports attached: None

This certificate applies only to the lot named above and describes the units sampled from it; it is not a guarantee for any other lot. Results were obtained by PrymaLab LLC's own laboratory (PrymaLab Analytical) and are not an independent third-party test unless an outside report is attached. This copy is issued for Puravium, a distributor of the lot; the brand shown does not change the laboratory, the method or the results. Reference values (CAS, formula, theoretical mass, sequence) are literature values for the named compound. Authenticity: scan the QR code or enter certificate number PL-26-00112 and key 3TW8THNN at prymalab.net/testing/verify; any copy that does not match the record shown there is not a PrymaLab certificate. PrymaLab LLC products are supplied for laboratory research use only and not for human or veterinary use.