

CERTIFICATE NO.

PL-26-00082

VERIFICATION KEY

LC7FVC2H

Scan or enter both at
prymalab.net/testing/verify

CERTIFICATE OF ANALYSIS · LOT RELEASE · ISSUED FOR PURAVIUM

Retatrutide 10 mg

Lot PL-2609-082 · Lyophilized powder · Label claim 10 mg/vial · Distributed as GIP3-RT

TESTED IN-HOUSE BY PRYMALAB

In-house analytical result,
not an independent third-party test

LOT / BATCH NUMBER	DATE MADE	DATE OF ANALYSIS	DATE ISSUED
PL-2609-082	2026-09-18	2026-09-19	2026-10-03
RETEST / EXPIRY	UNITS IN LOT / SAMPLED	STORAGE	CERTIFICATE TYPE
2028-09-18	400 / 3	-20 °C, desiccated, dark	Lot COA, 2 pages

COMPOUND	CAS NUMBER	MOLECULAR FORMULA	THEORETICAL MASS (AVG.)
Retatrutide	2381089-83-2	C ₂₂₁ H ₃₄₂ N ₄₆ O ₆₈	4731.4 Da
LY3437943; GIP/GLP-1/glucagon receptor triple agonist			
SEQUENCE			
H-Y-Aib-QGTFTSDYSI-[αMeLeu]-LDKK(AEEA-γGlu-C20 diacid)AQ-Aib-AFIEYLLEGGPSSGAPPPS-NH ₂ (39 residues; Aib2, α-methyl-Leu13, Aib20; Lys17 acylate d with eicosanedioic acid via γ-Glu and AEEA; C-terminal amide)			

PURITY · RP-HPLC, AREA %

99.24%

Specification 99.0 % • main peak at 8.62 min

IDENTITY · OBSERVED MASS

4731.1 Da

Theoretical 4731.4 Da • -0.3 Da (-63 ppm)

OVERALL CONCLUSION


PASS

Meets every specification for the tests performed

TEST	METHOD	SPECIFICATION	RESULT	CONCLUSION
Appearance	Visual inspection	White to off-white lyophilized powder	White lyophilized powder	PASS
Identity (mass)	LC-MS, ESI positive	Theoretical ± 1.0 Da	4731.1 Da	PASS
Purity	RP-HPLC, UV 214 nm, area%	99.0%	99.24 %	PASS
Largest impurity	RP-HPLC, area%	1.0%	0.25% at 7.44 min (Gln 3 /Gln 19 deamidation)	PASS
Water content	Karl Fischer	5.0%	1.7%	PASS
Bacterial endotoxin	LAL, kinetic	≤ 5.0 EU/mg	0.03 EU/mg	PASS
Sterility / bioburden	USP <71>, membrane filtration	No growth after 14 days	No growth; conforms	PASS

1 HPLC CHROMATOGRAM

Data file: RET_PL-2609-082_20260919_01 · Column C18, 150 × 4.6 mm · UV 214 nm



2 HPLC METHOD · RET_PL-2609-082_20260919_01

INSTRUMENT	DATA SYSTEM	DETECTOR
Waters Alliance e2695, SN E2695B742	Empower 3, Build 3471	Waters 2998 PDA
DETECTION WAVELENGTH	COLUMN	COLUMN PART / SERIAL
214 nm	C18, 150 × 4.6 mm, 3.5 µm	186003326 / USHX19700
MOBILE PHASE A	MOBILE PHASE B	FLOW RATE
0.1% TFA in water	0.1% TFA in MeCN	1.0 mL/min
COLUMN TEMPERATURE	INJECTION VOLUME	RUN TIME
40 °C	20 µL	12 min
SAMPLE DILUENT	SAMPLE CONCENTRATION	BLANK INJECTED
0.1% TFA in 30% MeCN	0.5 mg/mL	Yes
REFERENCE STANDARD		
In-house Retatrutide reference, lot PL-RS-RET-01, RT 8.59 min		

3 GRADIENT PROGRAM

TIME (MIN)	% A	% B
0.0	65	35
2.0	65	35
9.0	25	75
10.0	15	85
11.0	15	85
12.0	65	35

4 PEAK TABLE

PEAK	RT (MIN)	AREA	AREA %
1	6.01	5,677	0.12
2 (deamid.)	7.44	11,828	0.25
3	8.16	7,570	0.16
4	8.62	4,695,143	99.24
5	8.98	10,882	0.23
Total		4,731,100	100.00

Threshold 0.05% area · Blank subtracted: Y · Purity = main peak area ÷ total integrated area × 100, to two decimals; main peak shaded.

5 IDENTITY BY MASS SPECTROMETRY

INSTRUMENT	IONIZATION / MODE	OBSERVED M/Z (CHARGE)
Waters Acquity SQD	ESI, positive	947.2 (5+ [M+5H]) · 789.5 (6+ [M+6H])
DECONVOLUTED MASS	THEORETICAL MASS	DIFFERENCE
4731.1 Da	4731.4 Da	-0.3 Da (-63 ppm)
DATA FILE	ACCEPTANCE	
RET_PL-2609-082_MS_20260919_01	Within ± 1.0 Da	

6 ADDITIONAL TESTS

TEST	PERFORMED BY	DATE	REPORT / FILE
Water content	PrymaLab	2026-09-19	RPT-KF-PL-2609-082
Endotoxin	PrymaLab	2026-09-19	RPT-ET-PL-2609-082
Sterility	PrymaLab	2026-09-19	RPT-ST-PL-2609-082
Peptide content	PrymaLab	2026-09-19	RPT-AA-PL-2609-082

A test with no row here was not performed on this lot. Outside-laboratory reports, when used, are attached and named below.

7 REVIEW AND APPROVAL

D. Reynolds	S. Calloway	M. Phelps
Analyst · performed the testing	Reviewer · checked against raw data	Lab Director · approved for release
Date: 2026-09-19	Date: 2026-10-02	Date: 2026-10-03

Raw data retained at: Empower 3 (Build 3471), project PL-2609, archive PL-ARC-2026; 7 years. Outside laboratory reports attached: None

This certificate applies only to the lot named above and describes the units sampled from it; it is not a guarantee for any other lot. Results were obtained by PrymaLab LLC's own laboratory (PrymaLab Analytical) and are not an independent third-party test unless an outside report is attached. This copy is issued for Puravium, a distributor of the lot; the brand shown does not change the laboratory, the method or the results. Reference values (CAS, formula, theoretical mass, sequence) are literature values for the named compound. Authenticity: scan the QR code or enter certificate number PL-26-00082 and key LC7FVC2H at prymalab.net/testing/verify; any copy that does not match the record shown there is not a PrymaLab certificate. PrymaLab LLC products are supplied for laboratory research use only and not for human or veterinary use.