

CERTIFICATE NO.
PL-26-00064
VERIFICATION KEY
LX8XZZJL
Scan or enter both at
prymalab.net/testing/verify

CERTIFICATE OF ANALYSIS · LOT RELEASE · ISSUED FOR PURAVIUM

MOTS-c 40 mg

Lot PL-2609-064 · Lyophilized powder · Label claim 40 mg/vial

TESTED IN-HOUSE BY PRYMALAB

In-house analytical result,
not an independent third-party test

LOT / BATCH NUMBER PL-2609-064	DATE MADE 2026-09-05	DATE OF ANALYSIS 2026-09-07	DATE ISSUED 2026-10-03
RETEST / EXPIRY 2028-09-05	UNITS IN LOT / SAMPLED 250 / 2	STORAGE -20 °C, desiccated, dark	CERTIFICATE TYPE Lot COA, 2 pages


COMPOUND MOTS-c Mitochondrial open reading frame of the 12S rRNA type-c SEQUENCE H-MRWQEMGYIFYPRKLR-OH (16 residues; free N- and C-termini)	CAS NUMBER 1627580-64-6	MOLECULAR FORMULA C ₁₀₁ H ₁₅₂ N ₂₈ O ₂₂ S ₂	THEORETICAL MASS (AVG.) 2174.6 Da
---	----------------------------	---	--------------------------------------

PURITY · RP-HPLC, AREA %
99.56%
Specification 99.0% • main peak at 6.44 min

IDENTITY · OBSERVED MASS
2174.8 Da
Theoretical 2174.6 Da • +0.2 Da (92 ppm)

OVERALL CONCLUSION
PASS
Meets every specification for the tests performed

TEST	METHOD	SPECIFICATION	RESULT	CONCLUSION
Appearance	Visual inspection	White to off-white lyophilized powder	White lyophilized powder	PASS
Identity (mass)	LC-MS, ESI positive	Theoretical ± 1.0 Da	2174.8 Da	PASS
Purity	RP-HPLC, UV 214 nm, area%	99.0%	99.56%	PASS
Largest impurity	RP-HPLC, area%	1.0%	0.17% at 7.22 min	PASS
Water content	Karl Fischer	5.0%	1.7%	PASS
Bacterial endotoxin	LAL, kinetic	≤ 5.0 EU/mg	0.06 EU/mg	PASS
Sterility / bioburden	USP <71>, membrane filtration	No growth after 14 days	No growth; conforms	PASS



2 HPLC METHOD · MOTS_PL-2609-064_20260907_01

INSTRUMENT	DATA SYSTEM	DETECTOR
Waters Alliance e2695, SN E2695B742	Empower 3, Build 3471	Waters 2998 PDA
DETECTION WAVELENGTH	COLUMN	COLUMN PART / SERIAL
214 nm	C18, 150 × 4.6 mm, 3.5 µm	186003326 / USHX18211
MOBILE PHASE A	MOBILE PHASE B	FLOW RATE
0.1% TFA in water	0.1% TFA in MeCN	1.0 mL/min
COLUMN TEMPERATURE	INJECTION VOLUME	RUN TIME
30 °C	20 µL	10 min
SAMPLE DILUENT	SAMPLE CONCENTRATION	BLANK INJECTED
Water, 0.1% TFA	0.5 mg/mL	Yes
REFERENCE STANDARD		
In-house MOTS-c reference, lot PL-RS-MOT-01, RT 6.42 min		

3 GRADIENT PROGRAM

TIME (MIN)	% A	% B
0.0	95	5
3.0	95	5
7.0	55	45
8.0	20	80
9.0	20	80
10.0	95	5

4 PEAK TABLE

PEAK	RT (MIN)	AREA	AREA %
1	5.61	6,525	0.15
2	6.44	4,330,860	99.56
3	7.22	7,395	0.17
4	7.98	5,220	0.12
Total		4,350,000	100.00

Threshold 0.05% area · Blank subtracted: Y · Purity = main peak area ÷ total integrated area × 100, to two decimals; main peak shaded.

5 IDENTITY BY MASS SPECTROMETRY

INSTRUMENT	IONIZATION / MODE	OBSERVED M/Z (CHARGE)
Waters Acquity SQD	ESI, positive	1088.4 (2+ [M+2H]) · 725.9 (3+ [M+3H])
DECONVOLUTED MASS	THEORETICAL MASS	DIFFERENCE
2174.8 Da	2174.6 Da	+0.2 Da (92 ppm)
DATA FILE	ACCEPTANCE	
MOTS_PL-2609-064_MS_20260907_01	Within ± 1.0 Da	

6 ADDITIONAL TESTS

TEST	PERFORMED BY	DATE	REPORT / FILE
Water content	PrymaLab	2026-09-08	RPT-KF-PL-2609-064
Endotoxin	PrymaLab	2026-09-09	RPT-ET-PL-2609-064
Sterility	PrymaLab	2026-09-21	RPT-ST-PL-2609-064
Peptide content	PrymaLab	2026-09-08	RPT-AA-PL-2609-064

A test with no row here was not performed on this lot. Outside-laboratory reports, when used, are attached and named below.

7 REVIEW AND APPROVAL

D. Reynolds	S. Calloway	M. Phelps
Analyst · performed the testing	Reviewer · checked against raw data	Lab Director · approved for release
Date: 2026-09-07	Date: 2026-10-02	Date: 2026-10-03

Raw data retained at: Empower 3 (Build 3471), project PL-2609, archive PL-ARC-2026; 7 years. Outside laboratory reports attached: None

This certificate applies only to the lot named above and describes the units sampled from it; it is not a guarantee for any other lot. Results were obtained by PrymaLab LLC's own laboratory (PrymaLab Analytical) and are not an independent third-party test unless an outside report is attached. This copy is issued for Puravium, a distributor of the lot; the brand shown does not change the laboratory, the method or the results. Reference values (CAS, formula, theoretical mass, sequence) are literature values for the named compound. Authenticity: scan the QR code or enter certificate number PL-26-00064 and key LX8XZZJL at prymalab.net/testing/verify; any copy that does not match the record shown there is not a PrymaLab certificate. PrymaLab LLC products are supplied for laboratory research use only and not for human or veterinary use.